S.I.T./ISO 9001

PER pH - CONDUCIBILITÁ- REDOX

SOLUZIONI CERTIFICATE

Reagecon

I.S. EN ISO 9001

Reg. N°. 0012

La sempre maggior richiesta di soluzioni certificate, per la taratura di catene di misure formate da pHmetro – elettrodo pH o redox e conduttimetro – cella, trova una risposta in questo programma di tamponi per pH, Redox e standards per conducibilità, tutti forniti sempre con il certificato d'analisi S.I.T./ISO 9001.

Soluzione TAMPONE PROFESSIONAL +/- 0,010 pH

I tamponi professional sono pensati per i laboratori di ricerca che desiderano una precisione molto alta e vogliono scegliere tra una gamma di valori NBS.
Altri valori disponibili a richiesta.

12002083	1 x 500 ml	**pH 1,677 /20°C** incolore (NBS)
12002003	1 x 500 ml	**pH 4,001 /20°C** incolore (NBS)
12002013	1 x 500 ml	**pH 6,881 /20°C** incolore (NBS)
12002023	1 x 500 ml	**pH 7,429 /20°C** incolore (NBS)
12002033	1 x 500 ml	**pH 9,225 /20°C** incolore (NBS)
12002043	1 x 500 ml	**pH 10,062 /20°C** incolore (NBS)
12002053	1 x 500 ml	**pH 4,000 /20°C** (Rosso)
12002063	1 x 500 ml	**pH 7,000 /20°C** (Giallo)
12002073	1 x 500 ml	**pH 10,000 /20°C** (Blu)

ELETTROLITI RIEMPIMENTO E SOLUZIONI DI PULIZIA

Elettroliti di riempimento e soluzioni di pulizia per elettrodi pH, per migliorare l'efficienza e l'affidabilità dei vostri elettrodi.

12005033	1 x 100 ml	**1M LiCl in alcool etilico**

Soluzione STANDARD DI CONDUCIBILITÁ

Soluzioni standard di conducibilità per la taratura dei conduttimetri da laboratorio e da impianto. Ricordiamo che lo standard
µS ha una scadenza di 4 mesi, tutti gli altri 6 - 18 mesi.
Altri valori disponibili a richiesta.

12006233	1 x 500 ml	**10 µS/cm 25°C**
12006243	1 x 500 ml	**20 µS/cm 25°C**
12006273	1 x 500 ml	**200 µS/cm 25°C**
12006283	1 x 500 ml	**500 µS/cm 25°C**
12006293	1 x 500 ml	**1.000 µS/cm 25°C**
12006303	1 x 500 ml	**5.000 µS/cm 25°C**
12006343	1 x 500 ml	**10.000 µS/cm 25°C**
12006313	1 x 500 ml	**20.000 µS/cm 25°C**
12006323	1 x 500 ml	**50.000 µS/cm 25°C**
12006333	1 x 500 ml	**100.000 µS/cm 25°C**
12006383	1 x 500 ml	**150.000 µS/cm 25°C**
12006393	1 x 500 ml	**200.000 µS/cm 25°C**

SPECIALI

GIORGIO·BORMAC

Reagecon

Certificate of Quality

pH Buffer Solution

Buffer solution pH 7.00 ± 0.01 @ 20°C (Yellow)

Product No:	**10705C**
Lot No:	**72F1C**
Expiry date:	**28/06/04**

Mean pH value:	**7.00 @20°C**
Date of measurement:	**25/06/02**

pH measurement:
Tested in accordance with in-house method TPPHB. Measured with a combination glass electrode after multiple point calibration with reference materials.

Buffer substance:
Potassium di-hydrogen phosphate, Disodium hydrogen phosphate

Traceability:
Standard buffer solutions are certified traceable to the following U.S. National Institute of Standards and Technology

Standard Reference Materials:
NIST SRM 185g Potassium Hydrogen Phthalate, NIST SRM 186-I-e Potassium Dihydrogen Phosphate and NIST SRM 186-II-e Disodium Hydrogen Phosphate.

Date of issue of the certificate: 25/6/2002

Maeve Reddann

Quality manager: *Maeve Reddann*

Complementary information relative to this product is available at www.reagecon.com
Page 1 of 1. This Certificate must not be reproduced except in full.

TEMP. 11/02 REL. 1.3

DISTRIBUITO da:

GARANTITO da:

GIORGIO·BORMAC s.r.l.

GIORGIO·BORMAC